



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/593,591	06/13/2000	Gary K. Michelson	101.0078-00000	5619
22882 7590 12/17/2007 MARTIN & FERRARO, LLP 1557 LAKE O'PINES STREET, NE HARTVILLE, OH 44632			EXAMINER PREBILIC, PAUL B	
			ART UNIT 3774	PAPER NUMBER
			MAIL DATE 12/17/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

09/593,591

Applicant(s)

MICHELSON, GARY K.

Examiner

Paul B. Prebilit

Art Unit

3774

– The MAILING DATE of this communication appears on the cover sheet with the correspondence address –
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 06 June 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) See Continuation Sheet is/are pending in the application.
- 4a) Of the above claim(s) See Continuation Sheet is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) See Continuation Sheet is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____

Continuation of Disposition of Claims: Claims pending in the application are 1,2,4,5,7-20,22-30,32-44,46,47,49-86,88,89,91-129,148-164,166,168-184,186,188-204,206,208-224,226,228-264,266 and 268-282.

Continuation of Disposition of Claims: Claims withdrawn from consideration are 43,44,46,47,49-84,153,154-157,183,184,186,188-202 and 243-262.

Continuation of Disposition of Claims: Claims rejected are 1,2,4,5,7-20,22-30,32-42,85,86,88,89,91-129,148-152,155-164,166,168-182,203,204,206,208-224,226,228-242,263,264,266 and 268-282.

Election/Restrictions

Claims 43, 44, 46, 47, 49-84, 153, 154-157, 183, 184, 186, 188-202, and 243-244, 246, 248-262 are withdrawn from further consideration pursuant to 37 CFR 1.142(b) as being drawn to a nonelected species, there being no allowable generic or linking claim. Election was made **without** traverse in the reply filed on January 26, 2005.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1, 2, 4, 5, 7-12, 14, 16, 85, 86, 88, 89, 91-96, 98, 100, 102-105, 116, 117, 163, 164, 166, 168-173, 175, 179-180, 203, 204, 206, 208, 209, 211, 213, 215, 223, 224, 228-229, 231, 233, 235, 239, 240, 263, 264, 266, 268-269, 271, 273, 275, 279, and 280 are rejected under 35 U.S.C. 102(e) as being anticipated by Coates et al (US 5,989,289). Coates anticipates the claim language the bone ring as claimed is obtained from a long bone (see column 11, lines 42-61), the leading end as claimed is the posterior wall (115), and the trailing end is the anterior wall (111); see Figures 1 to 4.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-2, 4, 5, 7-20, 22, 25, 30, 32-35, 85, 86, 88, 89, 91-106, 109, 114-119, 127-129, 148, 158, 163, 164, 171, 174, 175, 179, 180, 203, 204, 206, 208-211, 213-215, 219, 220, 223, 224, 226, 228-231, 233-235, 239, 240, 263, 264, 266, 268-271, 273-275, 279, and 280 are rejected under 35 U.S.C. 103(a) as being unpatentable over Boyle et al (US 6,277,149) alone. Boyle discloses a bone ring obtained from a long bone (Figures 5, 10, 12, and column 2, lines 28-55) that has all the claimed structure except for non-arcuate upper and lower surfaces as claimed. However, since Boyle teaches in other embodiments that it was known to utilize non-arcuate upper and lower surfaces (see Figures 1, 3, 9, 12, and 13), it is the Examiner's position that it would have been obvious to make the Figure 19 embodiment with non-arcuate upper and lower surfaces for the same reasons that Boyle utilizes the same in the other embodiments

Regarding claims 25 and 109, the plural openings are the pores in the bone that are naturally present in such material.

Claims 26-29, 110-113, 149-152, 159-162, 172, 177, 178, 182, 212, 217, 218, 222, 232, 237, 238, 242, 272, 277, 278, and 282 are rejected under 35 U.S.C. 103(a) as being unpatentable over Boyle et al (US 6,277,149) alone as used in the previous rejection, further in view of Kozak et al (US 5,397,364). Boyle fails to disclose the screw holes and screws as claimed. However, Kozak teaches that it was known to put multiple bone screws in similar implants as a means to secure them; see Figures 24 and 25 and the textual description of these figures. Therefore, it is the Examiner's position that it would have been obvious to use bone screws and have holes for them in the Boyle implant for the same reasons that Kozak uses the same: to secure the implant to the adjacent bone.

With regard to claims 28 and 29, it is the Examiner's position that it would have been prima facie obvious to make the bone screws and other bone implant components out of bone because the ring itself is made of that material. Since the ring itself is made of bone material, it would have been obvious to make the other parts of the implant out the same material because it would have all the advantages and purposes that led Boyle to make the bone ring out of bone or composites of bone.

With regard to claim 155, it is the Examiner's position that it would have been prima facie obvious to have more than two bone screw openings since both one and two bone screws and their openings are taught by Boyle.

Claims 36-42, 120-126, 166, 168-170, 173, 181, 206, 221, 241, and 281 are rejected under 35 U.S.C. 103(a) as being unpatentable over Boyle et al alone as applied in the rejection of claim 1 above in further view of the suggestions of Boyle (US


6,277,149). Boyle et al fails to disclose the types of growth promoting substances claimed. However, the Examiner posits that the type of growth promoting substance would have been prima facie obvious over Boyle et al because there is no criticality for such types shown or demonstrated. For example, in column 4, lines 10-13, Boyle discloses putting bone growth material into the opening (26). The specific type of bone implant material claimed would amount to a substitution of a functionally equivalent bone growth material for another one as disclosed by Boyle. For this reason, it would have been prima facie obvious to an ordinary artisan to substitute one bone growth material for another since the other one would function equally well.

Claims 23, 24, 107, 108, 176, 216, 236, and 276 are rejected under 35 U.S.C. 103(a) as being unpatentable over Boyle et al (US 6,277,149) in view of Zdeblick et al (US 5,669,909). Boyle et al renders obvious the claim language as explained in the rejection of claim 1 above but fails to disclose more than one bore through the upper and lower surfaces as claimed. Zdeblick et al, however, teaches that it was known to have more than one hole in similar implants; see the figures. Hence, it is the Examiner's position that it would have been obvious to have two or more holes in the Boyle et al implant for the same reasons that Zdeblick et al has the same and in order to promote greater ingrowth.


Response to Arguments

Applicant's arguments filed October 16, 2007 have been fully considered but they are not persuasive.

The Applicant argues that Coates does not disclose a majority of the opening being between the perpendicular plane and the leading end. However, as illustrated below, the Examiner asserts that the claim language is fully met in this regard because the perpendicular plane can be anywhere along the length of the device as long as it bisects the length. "Bisecting" is understood to mean divide into two parts since that is the broadest ordinary meaning of the term; see Encarta Online Dictionary at <http://encarta.msn.com/encnet/features/dictionary/DictionaryResults.aspx?refid=1861590800> that defines "bisect" as "split into two: to split something into two parts." It is noted that the term "bisect" or "bisecting" is not used in the present specification other than in the claims. For this reason, the rejections utilizing Boyle have been maintained as well.



The rejections utilizing Boyle were traversed in a similar manner. Again the highlighted area is the majority of the area of the opening as claimed.



Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

Applicant should specifically point out the support for any amendments made to the disclosure, including the claims (MPEP 714.02 and 2163.06). Due to the procedure outlined in MPEP 2163.06 for interpreting claims, it is noted that other art may be

Application/Control Number:
09/593,591
Art Unit: 3774

Page 9

applicable under 35 USC 102 of 35 USC 103(a) once the aforementioned issue(s) is/are addressed.

Applicant is respectfully requested to provide a list of all copending applications that set forth similar subject matter to the present claims. A copy of such copending claims is respectfully requested in response to this Office action if the application is not stored in image format (i.e. the IFW system) or published.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Examiner Paul B. Prebilic whose telephone number is (571) 272-4758. He can normally be reached on 6:30-5:00 M-Th.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on 571-272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

/Paul Prebilic/
Paul Prebilic
Primary Examiner
Art Unit 3774